



Ministerie van Volksgezondheid,
Welzijn en Sport

Veldbijeenkomst implementatie MDR/IVDR

27 november 2019



Programma

1. Best practices door het veld
 - Verschillende voorbeelden van zowel fabrikanten als zorginstellingen
2. Nationale ontwikkelingen
 - Update wetgeving
 - Update CCMO

Pauze

3. Europese ontwikkelingen
4. Vraag en antwoord

Afsluiting en borrel



- Best practices door het veld
 1. *Nefemed en Firevaned*
 2. *FHI en FME*
 3. *NFU en NVZ*



Best practices door het veld

- Firevaned en Nefemed
- Patrick Bakker – Zimmerbiomet
- Leander Leijh - Medux




Best practices door het veld: MDR

Patrick Bakker
QA/RA Specialist
Zimmer Biomet
Nederland B.V.




*Ons inzetten voor de
hoogste norm van
patiëntveiligheid, kwaliteit
en integriteit*

We zetten ons in voor de hoogste norm van patiëntveiligheid en kwaliteit in onze producten en diensten en erkenning voor integriteit en ethisch zaken doen.



Interne puzzel





Quarterly Update



Quarter 1

- Establish Steering Committee
- Create & Approve Charter
- Onboard Dedicated Team Members
- Initiate Internal Communication Plan
- Initiate External Communication Plan

Quarter 2

- Scope Product – 59%
- Scope Regulation – Continuous Process
- Create all Workstream Sub-Charters w/ Project plans
- Approve all Workstream Sub-Charters w/ Project plans

Quarter 3

- Conduct Workstream Gap Assessments
- Create Workstream Implementation Plans
- Finalize Workstream Budgets

Quarter 4

- Executing Implementation Plan

Completed
 In Progress
 In progress but delayed
 In progress , delay risk
 Not Started



End of Transition
Period In... **2** **10** **13**
YY MM DD

Budget
Deadline In... **2** **13**
MM DD

14

Veranderen





Implantaatkaart (Artikel 18 MDR)

Patient Card Draft

<p>NAME: CLS* Spotorno* MODEL: 29.00.09-050 LOT: 999999999 UDI-DI: 00889024394407</p> <p>Zimmer GmbH, Sulzerallee 8, 8404 Winterthur, Switzerland</p> <hr/> <p>NAME: Continuum* Trabecular Metal™ MODEL: 00-8757-046-00 LOT: 999999999 UDI-DI: 00889024151277</p> <p>Zimmer, 1800 W Center St, Warsaw, IN USA</p> <hr/> <p>Affix Implant Label</p> <hr/> <p>Affix Implant Label</p> <hr/> <p>Affix Implant Label</p> <hr/> <p>Affix Implant Label</p>	<p>99-8888-123-EN Rev 1 / ©2018</p> <p>For additional implant information go to website and enter MODEL or UDI number. Website: www.zimmerbiomet.com/implant Alternatively, call the Zimmer Biomet Customer Service at: Phone: +41 (0)58 854 80 00</p> <p> ZIMMER BIOMET</p> <p>Zimmer GmbH Sulzerallee 8 CH-8404 Winterthur Switzerland</p> <p>Patient Implant Card</p> <hr/> <p>Patient name</p> <hr/> <p>Doctor/Hospital information</p>
---	---



Enkele punten als voorbeeld:

- e-IFU
- Hogere mate van samenwerking met onze toeleveranciers
- Afnemers/ziekenhuizen sterkere ketenpartner
- door MDR wordt er in de transitie al steeds meer dezelfde taal gesproken, partijen begrijpen elkaar beter.




Bedankt voor uw aandacht!

Patrick Bakker
QA/RA Specialist
Zimmer Biomet Nederland B.V.
patrick.bakker@zimmerbiomet.com
+31 6 2395 6414



ZIMMER BIOMET
Your progress. Our promise.®



medux

Implementatie MDR

25-11-2019, Leander Leijh



meer
mogelijk
maken.

Inhoud

- Voorbereiding op de MDR
- Risico's en kansen
- Waar staan we nu
- What's next



Vorbereiding op de MDR

- Vaststellen van de belangrijkste wijzigingen
- Benoemen van de thema's die invloed kunnen hebben vanuit onze rol
- Deelnemen in branche overleggen
- Vaststellen van risico's en op zoek naar kansen




- Grootste risico's:
 - Verhoging van administratieve lasten
 - Niet kunnen continueren bepaalde producten
 - Onvoldoende aandacht voor verschillen tussen risicoklassen
 - Toename in kosten
- Kansen:
 - Toonaangevend zijn door in te spelen op de risico's
 - Versterken samenwerking in de keten
 - Optimaliseren kwaliteit en daarmee kosten reduceren



Waar staan we nu?


- Nog niet alles is bekend, maar veel al wel
- Fit-Gap analyse op processen
- Dialoog met toeleveranciers
- Afstemming met klanten; wat leeft daar?
- In samenspraak met branche kijken naar mogelijkheden om zaken te vereenvoudigen



What's next?

- Veel moet nog komen en geconcretiseerd worden in de keten
- Tot medewerker niveau uitleggen wat de impact op werk is
- Doorrekenen gevolgen en vaststellen financiële impact





medux

Medux B.V.

Reactorweg 160
3542 AD Utrecht

www.medux.nl

meer
mogelijk
maken.



Best practices door het veld

- FHI en FME
- Rick Paauw - Medtronic

EU MDR

VELDBIJEENKOMST

@ VWS

November 27 2019



Medtronic

MDR IMPLEMENTATION ACTIVITIES AT MEDTRONIC



We analyzed the changes and the impact to our company and products. We then took action to address the implementation across the company.




>76,000 models (CFNs)*
> 95,000 barcodes (GTINs)*
>2/3 of products sold outside EU as well

*Based on SAP report 11/29/2016; adjustment needed to remove obsolete materials and associated information.



>1530 technical documents
>750 CERs, 13 Notified Bodies
> 97,000 package labels
>140 quality systems covering
>65,000 quality documents

Article Chapter	Key Regulated Elements	Related Annexes
Scope & Definitions	Devices with non-viable human tissues, Devices for aesthetic purposes, Definitions of economic operators	Annex XVI
Making Available of Medical Devices	Common Specifications, Economic Operators, SUD Reprocessing, PRR, Implant Cards, DoC, System & Procedure Packs	Annexes I, IV, V, XII
Identification and Traceability	UDI Database Structure and Data Elements, UDI Definitions, Registration of Economic Operators & Devices, SSCP	Annex VI
Notified Bodies	Notified Body Qualification Requirements, Designation Process and Auditing by Competent Authorities	Annex VII
Classification & Conformity Assessment	Up-classifications, Conformity Assessment Procedures, Panel Review and Scrutiny, Safety and Performance Requirements	Annexes I, II, VIII, IX, X, XI, XIII
Clinical Evaluation & Investigation	Restriction of Equivalence Reference, Clinical Investigation Application, Conduct and Reporting	Annexes II, XIV & XV
Post-Market Surveillance	Post-Market Plan, Periodic Safety Report, Incident and Trend Reporting (Vigilance), Market Surveillance by CAs	Annex III



CHANGES TO VIGILANCE REPORTING CRITERIA



We mapped the regulatory requirements by data of application.

Regulatory Requirements by DoA	MDR			
	Class I (nm,ns,nr)	Class I <u>r</u> <u>si</u>	Class I <u>(s,m)</u> , Class IIa, IIb	Class III, IIb implantables
QMS recertification for MDR				
NB QMS/Tech Doc review for MDR				
Economic Operator definition/contracting				
Technical Documentation (incl GSPR)				
DoC updates				
CE registration in Eudamed				
Economic operator registration				


CHANGES TO VIGILANCE REPORTING CRITERIA



Reporting will increase, based on historical data and MDR requirements.

MEDDEV 2.12 rev 8 Vigilance Reporting	Current MDD/AIMDD	EU MDR
5.1.7 Report immediately, but no later than 2 days	Reportable	Reportable (no change)
5.1.7 Report immediately, but no later than 10 days	Reportable	Reportable (no change)
----- Report immediately, but no later than 15 days (CZ)	Reportable	Reportable (no change)
5.1.7 Report immediately but no later than 30 days	Reportable -> 30 days	Reportable -> 15 DAYS (CHANGE)
5.1.2 Periodic summary reporting	Reportable	Reportable (conditions&format changed)
5.1.1.B No evidence event is product related and/or device tested satisfactorily	Not reportable	Not Reportable 15 day deadline may impact
5.1.1.C No risk of death or serious injury	Not reportable	Not Reportable (no Change)
5.1.3.1 Deficiency of a device found by user prior to its use	Not Reportable	Reportable (CHANGE)
5.1.3.2 Event caused by patient conditions	Not Reportable	Reportable (CHANGE)
-----	Not Reportable	Reportable (CHANGE)
-----	Not Reportable	Reportable (CHANGE)

ECONOMIC OPERATORS WITHIN MEDTRONIC




Documentation on/Information accompanying the device

Each actor in the supply chain checks the compliance of the previous one

<p>Manufacturer</p> <ul style="list-style-type: none"> • Draw up technical doc, DoC and affixes CE mark • UDI carrier on the device label and higher levels of packaging • IFU in an official EU language • Label correct 	<p>Authorised rep</p> <ul style="list-style-type: none"> + Verify technical doc and conf. assessment + Keep a copy of the technical doc, DoC, certificate 	<p>Importer</p> <ul style="list-style-type: none"> + Verify that a manufacturer is identified and AR is designated + Importer details on device label or on its packaging or in an accompanying doc 	<p>Distributor verify:</p> <ul style="list-style-type: none"> • CE mark and EU DoC • Label and IFU • UDI
--	---	---	--

TRAINING OF ALL EMPLOYEES ON EU MDR

Face-to-face sessions



eLearning and WebEx sessions



GUIDANCE AND INTERPRETATION



We developed a **playbook** We conducted a gap analysis

We monitor changes and make updates

EU Medical Device Regulation (EU MDR)
Medtronic Guidance Playbook: System Governance
 Based on EU MDR text published May 05, 2017 in the Official Journal of the European Union
 Version 1, August 7, 2018
 Author: Joachim Wilke

SCOPE: ARTICLES 1, 2, 4 AND ANNEX XVI

1. Introduction

The EU MDR applies to products for human use which fulfill the definition for medical devices (Article 2 (1)) or their accessories (Article 2 (2)) and are intended to be placed on the EU market. Details on the EU MDR scope and respective product qualification processes in the EU are in particular regulated in Article 1 and 4 of the Regulation.

The EU market should be understood as the territory of the Member States of the European Union amended by the countries of the European Free Trade Association (EFTA), i.e., Norway, Iceland and Liechtenstein. In addition, Switzerland intends to conclude with the EU on a Mutual Recognition Agreement (MRA) which allows placing products which are compliant with EU MDR to be placed on the Swiss market. Note that such a MRA is already in place for the AIMDD/MDR but needs to be renewed for the EU MDR.

In comparison to the AIMDD/MDR, the scope of the EU MDR has been extended to certain product groups listed in Annex XVI with a non-medical purpose. Also, medical devices which are manufactured utilising tissues or cells of human origin, or their derivatives, which are non-viable or are rendered non-viable (Article 1 (6) (g)), or incorporate as an integral part non-viable tissues or cells of human origin or their derivatives that have an action ancillary to that of the device (Article 1 (10)), have been included in scope of the EU MDR.

The EU MDR applies also to devices including their accessories used in clinical investigations conducted in the EU.

2. Impact Summary

Although several sections of the AIMDD/MDR have been modified in the EU MDR text, the impact on the current Medtronic portfolio is considered to be low. There is no Medtronic product which has been classified as being in scope of the AIMDD/MDR which is not in scope of the EU MDR. This is in particular true for products that contain or consist of viable biological material or viable organisms, including living micro-organisms, bacteria, fungi or viruses, in order to achieve or support the intended purpose of the products which are now not anymore classified as medical devices. Further, Medtronic does not place products as listed in Annex XVI with a non-medical use on the EU market. The EU MDR scope extension to devices which are manufactured utilising tissues or cells of human origin, or their derivatives, which are non-viable or are rendered non-

EU Medical Device Regulation (EU MDR)
AIMDD / MDD vs. EU MDR Gap Analysis – Articles
 Based on EU MDR text published May 05, 2017 in the Official Journal of the European Union

Medtronic confidential – for Internal use only
 Version 4, updated Dec 30, 2018
 Author: Joachim Wilke

Note:

- Abbreviations are spelled out on first reference and in the EU MDR abbreviations document posted on the Sitebuilder.
- Text in red indicates changes between the final EU MDR regulation text published May 5, 2017 and the previous version (Consolidated Trilogue Compromise Text issued June 27, 2016)

Note: This document tracks changes between versions 3 and 4. If you wish to see it without the edits, navigate to the Review tab and view it in Final (or "No Markup") mode. If you wish to see the edits, view it in "All Markup" mode.

Row	AIMDD 90/385/EEC	MDD 93/42/EEC	EU MDR Text Regulation EU 2017/745 (May 5, 2017)	Comments
1.			Chapter 1 Scope and definitions	
2.			Article 1 Scope	Changes to the scope of the EU medical device legislation compared to the Active Implantable Medical Device Directive (AIMDD) and Medical Device Directive (MDD) as outlined in this article may require an update of existing quality management system (QMS) documents.
3.	Member States shall take all necessary steps to ensure that the devices may be placed on the market and/or put into service only if they comply with the requirements laid down in this Directive when duly supplied, properly implanted and/or properly installed, maintained and used in	Member States shall take all necessary steps to ensure that devices may be placed on the market and/or put into service only if they comply with the requirements laid down in this Directive when duly supplied and properly installed, maintained and used in accordance with their	1. This Regulation lays down rules concerning the placing on the market, making available on the market or putting into service of medical devices for human use and accessories for such devices in the Union. This Regulation also applies to clinical investigations concerning such medical devices and accessories conducted in the Union.	The EU MDR scope refers not only to medical devices and their accessories, but also to all processes linked to the requirements for placing of such products on the EU market, including clinical investigations.

EU MEDICAL DEVICE REGULATION NO. 745/2017 (EU MDR)

Mini-Playbook

A compilation of the first three sections of Medtronic's EU MDR Playbook Guidance Chapters

Updated January 09, 2019



FME



**POWERED
BY DUTCH
TECHNOLOGY**

FME Zorg

FME ZORG


EU MEDICAL DEVICE REGULATION (EU) 2017/745

MDR CHALLENGES

- Notified Body designation
- Drug device combination consultation
- Grace period
- PMCF
- MDR product information requests from customer
- Eudamed (2 years postponed)
- Local Laws

FME ZORG

EU MEDICAL DEVICE REGULATIONS (EU) 2017/745 NOTIFIED BODY DESIGNATION



Medical Device recertifications delay

Currently thousands of life saving/changing devices are waiting at the desk of the Certification Bodies waiting for certificates to be issued. These are piling up since CB's don't have the resources to issue the certificates prior May 2020.

Proven and/or essential / life-saving technologies maybe interrupted / not available due to a (currently perceived) capacity issue at notified bodies – to issue renewal certificates under the current guidance (MDD/AIMD – derogation till May 2024)

State of the art definition discussion

Some of the above referenced products, may not meet the “state of the art” definition – should have been covered under AIMD/MDD Certification – the real issue may be the absence of additional or “sufficient” clinical data, as required per the MDR. Although they have been in the market for years without patient safety issues or concerns.

Smaller companies discontinue because they don't have the resources/budget. This might contribute to more product disruption / shortage / total unavailability of certain products.

FME ZORG

EU MEDICAL DEVICE REGULATION (EU) 2017/745 DRUG DEVICE COMBINATION CONSULTATION



Device-drug combination products: Shall the 210-day pharma authority consultation be repeated for legacy products?

10 May 2019

MedTech Europe is seeking a clarification on the appropriate procedure to be applied by Notified bodies in upgrading devices compliant to Directives 93/42 on Medical Devices (MDD) and 90/385 on Active Implantable Medical Devices (AIMD), which incorporate a substance which could be considered a medicinal product, to the requirements of Regulation 2017/745 on Medical Devices (MDR).

MedTech Europe's interpretation of Section 5.2 of Annex IX of the Regulation 2017/745 is that, in principle, there will not be the need to require again a scientific opinion from one of the competent bodies established in ... Directive 2001/83 on Medicinal Products for Human Use or with ... medicinal substance and the way it is

- FME is aligned with the MedTech position paper regarding:

Device-drug combination products and if the 210-day pharma authority consultation shall be repeated for legacy products

FME ZORG

EU MEDICAL DEVICE REGULATION (EU) 2017/745 DRUG DEVICE COMBINATION CONSULTATION

Section 5.2 (a) of Annex IX of the Regulation 2017/745

Devices that incorporates

- **as an integral part** , - **a substance**, - **which, if used separately**


may be considered to be a medicinal product within the meaning of Dir. 2001/83/EC

Section 5.2 (b)

The NB shall, seek a scientific opinion from a CA designated in accordance with Directive 2001/83/EC or from the EMA...

Section 5.2 (f)

- Before any change is made with respect to an ancillary substance,....., the manufacturer shall inform the NB.
- The NB shall seek the opinion of the medicinal products authority, in order to confirm that the quality and safety of the ancillary substance remain unchanged.



FME ZORG

EU MEDICAL DEVICE REGULATION (EU) 2017/745 **DRUG DEVICE COMBINATION CONSULTATION**

Interpretation of the AIMD, MDD and MDR, is that there is no change to the rationale, or the scope for a scientific consultation.

Therefor it could be possible for a NB to evaluate the existing data, to determine whether the scientific opinion, independently under which legal regime it was delivered, is positive and in place.

This should be possible as long as there is no change in:

- the medicinal substance
- the manufacturing process
- and the way its integrated with the device remain the same

To comply with the MDR it's our interpretation of section 5.2 of annex IX, that a scientific opinion in accordance with Directive 2001/83 is not required again

This will avoid, a repetition of an unchanged Pharma Authority consultation with a potential delay up to 210 days

FME ZORG

EU MEDICAL DEVICE REGULATION (EU) 2017/745 GRACE PERIOD

The European Commission is considering a corrigendum for extending the grace period for class I products - specifically, class I reusable surgical instruments and up-classified class I devices that require a notified body review by the date of application (DoA).

This would allow four additional years to comply and allow more time to focus on higher-risk devices.

It means an additional Grace period for:

- Class Ir
- Up-classified class I devices that require a notified body review by the (DoA)
Stand-alone SW could also benefit from the corrigendum if up-classified by MDR, (rule 11)

What about ?

- Class I products/software that become or that stay class I - Most likely not covered
 - New products previously not falling under the MDD/AIMD - No, not covered
- For Annex XVI implementing acts may contain a grace period, unclear yet

FME ZORG

EU MEDICAL DEVICE REGULATION (EU) 2017/745 POST MARKET CLINICAL FOLLOW-UP

Article 61 (11): The clinical evaluation and its documentation shall be updated throughout the life cycle of the device concerned with clinical data obtained from the implementation of the manufacturer's PMCF plan

Clinical centers foresee challenges to meet the manufacturers expectations because:

- # of registries will likely increase
- Hospital will get a large amount of request got from Manufacturers to collect & share clinical data
- Hospitals don't have the budget / resources to facilitate these requests
- Hospitals have a lot of data but this can't easily pulled from the systems
- Providing manufacturers access to the hospitals system is difficult from GDPR / AVG point of view
-

STAKEHOLDERS NEED TO WORK TOGETHER TO UNDERSTAND EACH OTHERS.....,

FME ZORG

EU MEDICAL DEVICE REGULATION (EU) 2017/745 MDR PRODUCT INFORMATION REQUEST

A	B	C	D	E	F	G	H	I	J	K	L	M	N	O	P	Q	R	S	T	U	V	W	X	Y	Z	AA	AB	AC	AD	AE	AF	AG	AH	AI	AJ
1. Article no.	2. Z-Index nummer (Zicode)	3. Supplier article no.	4. Article name / description	5. Ref number	6. Intended Use Article	7. CE-mark	8. Conform regulations	9. Certificate code (CE certificate or Declaration of conformity ID)	10. Certificate expiry date	11. Current risk class MDD/IVD	12. New risk class MDR/IVDR	13. Notified Body no.	14. Is the medical device covered by a clause in Art 12 MDD 93/4	15. Name Manufacturer	16. European representative when manufacturer is not EU	17. Piece UDI (EAN)	18. Inner carton EAN	19. Outer carton EAN	20. Pallet amount	21. Single use	22. Is education before processing for re-use of the product necessary?	23. Is the item sterile?	24. If not sterile: is sterilisation possible? If yes, please fill out attachment	25. Is training for safe use required?	26. Specific storage conditions (indicate temp.)	27. Are there specific requirements regarding waste processing of the product?	28. Article contains latex	29. Article contains phthalates	30. Article contains Bisphenol A	31. The product is made using nanotechnology	32. If item is supplied sterile, is the shelf life less than 24 months after sterilization?	33. Attachment name (Specification document for Steril. Products)	34. Attachment name (CFR Declaration of conformity)	35. Attachment name (Declaration of conformity)	36. Attachment name (IFU (Instruction For Use))

STATEMENT:

The new EU Medical Device Regulation (MDR) is welcomed by Medtronic as it will strengthen the system for "CE marking" of medical devices and post-market surveillance of CE marked devices. Medtronic is committed to ensuring the quality, safety and performance of its products and therapies. We will continue to excel in meeting the strengthened standards and goals of the MDR, and as we transition the certification of our products.

New Regulation

- The European Union has adopted and published a new regulation related to medical devices.
- This new regulation strengthens the current system for "CE marking" of medical devices, including new requirements for post-market surveillance.
- The new regulation begins to apply as of May 26, 2020. The regulation provides for continuity from the current to the enhanced rules in the form of transition provisions, meaning that products with valid certificates under the current directive(s) can be made available until 2025 as long as they additionally meet a certain number of the MDR requirements such as post-market surveillance.
- The new regulation introduces several important improvements to modernize the current system, including control of manufacturers, importers, authorized representatives and distributors, increased requirements for clinical evidence and technical documentation, improvements related to risk management and vigilance reporting, an EU database for important information related to medical devices, and overall increased transparency.

What's changed?

- Medical Devices are transitioning from being CE marked under the two current (and separate) pieces of legislation, the Medical Devices Directive and the Active Implantable Medical Devices Directive, to being CE marked under a single new Medical Device regulation.
- In general, the regulation retains all the requirements of the Directives, while adding some new ones. Compared to the current Directives, the new regulation emphasizes a life-cycle approach to safety, a reinforced emphasis on clinical data and increased transparency via a publicly accessible European database on devices.

Product availability and transition timelines

- The MDR will apply from the 26th May 2020. With many thousands of products to be then assessed under the revised system from that date, the MDR allows a gradual phasing in to avoid market disruption and to ease the transition from Directives to a Regulation.
- The system provides a transition for the following to happen:
 - Applications for the many thousands of devices currently on the market under the Directives are submitted by manufacturers to be reviewed under the Regulation
 - The Notified Bodies review these applications including assessment of the data and assessments of the individual design and manufacturing sites.
 - During that time:
 - Products with effective certificates under the current directives can be placed on the market until May 2024
 - one additional year is allowed for products already placed on the market to be made available (e.g., by a distributor).
 - After May 2025, products certified under the current directive(s) can no longer be made available on the EU market.
 - During the transition, EU Member State authorities and Notified Bodies will continue to oversee the EU market, ensure the medical devices are safe and perform as intended, regardless of whether they are CE marked under the current Directives or the new Regulation.
 - Medtronic has created a comprehensive implementation plan to ensure that our products meet the certification requirements of the new regulation.
 - Regulatory documentation such as Declarations of Conformity, certificates, labels and instructions for use issued under the current Directives, remain valid until transition is complete, and can be used in regulatory submissions and tenders.

A	B	C	D	E	F	G	H	I	J
Veld-/Field: Rood verplicht/Orange verplicht indien van toepassing Red mandatory/Orange mandatory if applicable	ZXL Artikelnummer ZXL Tradetitem ID	ZXL ArtikelOmschrijving ZXL Tradetitem description	ZXL Leverancier nummer ZXL Supplier number	Artikelnummer Tradetitem ID	Artikelnummer Fabrikant (PCN/REF) Manufacturer part number (MPN/REF)	MDR relevant (Y/N) MDR relevant (Y/N)	UOMkleinsteBasisEenheid UOMSmallestBaseUnit	Barcode BasisEenheid GTIN Barcode BaseUnit GTIN	Barcode BasisEenheid HIBC Barcode BaseUnit HIBC
Vervul de in gerespecteerd input:	n.v.t./not applicable	n.v.t./not applicable	n.v.t./not applicable	n.v.t./not applicable	n.v.t./not applicable	Boolean	Tekst/Text	Numeriek/Numeric	Tekst/Text
UOM BinnendoosEenheid UOM InnerboxUnit	Barcode BinnendoosEenheid GTIN Barcode InnerboxUnit GTIN	Barcode BinnendoosEenheid HIBC Barcode InnerboxUnit HIBC	UOM Besteleenheid UOM OrderUnit	Barcode Besteleenheid GTIN Barcode OrderUnit GTIN	Barcode Besteleenheid HIBC Barcode OrderUnit HIBC	Risico-klasse medische producten (Huidia) Risk-class medical tradetitems (Current)	Risico-klasse medische producten (Nieuw) Risk-class medical tradetitems (New)		
Tekst/Text	Numeriek/Numeric	Tekst/Text	Tekst/Text	Numeriek/Numeric	Tekst/Text	Tekst/Text	Tekst/Text		

STAKEHOLDERS NEED TO WORK TOGETHER TO UNDERSTAND EACH OTHERS

FME ZORG

EU MEDICAL DEVICE REGULATION (EU) 2017/745 EUDAMED DELAYED

Eudamed - Database intended to increase the transparency of the medical device regulatory system.

Certain parts will be available to the public:

- Economic operators
- Devices registered
- Certificates
- Field safety notices
- Parts of vigilance reports
- Clinical study protocols/reports
- Summary of Safety and Clinical Performance (SSCP)

EUDAMED DELAYED WITH 2 YEARS
MCDG guidance expected Dec/Jan timeframe

FME ZORG

EU MEDICAL DEVICE REGULATION (EU) 2017/745 LOCAL LAWS

What local laws (derogation laws) are expected before DOA.

IGJ is working on a guidance document regarding PMS reporting.

- What else in the pipeline
- What is the role of TA's
- Collaboration is key to guarantee continuity of care



Best practices door het veld

- NFU en NVZ
- Astrid Verkaar - NVZ

Veldbijeenkomst MDR/IVDR

Astrid Verkaar

3 december 2019



Nederlandse
Vereniging van
Ziekenhuizen

Implementatie MDR en IVDR in ziekenhuizen

- / Voldoen aan wet- en regelgeving
- / Noodzakelijk: duiding en toetsingskader om optimaal te kunnen werken
- / Wetgeving mag veilige patiëntenzorg niet in de weg zitten



Rol NVZ en NFU



NFU: programma “Veilige patiëntenzorg door veilige technologie”

/ Belangrijkste veranderingen zijn in kaart gebracht (17)

/ Expertteams geformeerd:

/ Beschikbaarheid

/ Traceerbaarheid

/ Klinisch Onderzoek


/ Medische Hulpmiddelen

De veranderingen 1/2

#	Verandering	MOR en/of IVDR	Belanghebbende	Uitvoering door	Wat moet er gebeuren?
V1	Alle Medische hulpmiddelen (inclusief klasse III-implantaten) krijgen een unieke identificatiecode zodat traceerbaarheid voor de patiënt geborgd is. Daar bovenop is voor klasse III-implantaten de verplichting om de implantaatgegevens te registreren zodat ze aan de patiënt kunnen worden verstrekt samen met de implantaatkaart.	MOR	Patiënt	Expertteam Traceerbaarheid	De inclusielijst wordt uitgebreid met alle klasse III implantaten. Huidige implementatie van het LIR dient te worden uitgebreid met de klasse III implantaten. Er is dus geen nieuw (elektronisch) systeem nodig. Onderzoeken welke voordelen dit kan bieden voor supply chain management.
V2	De patiënt krijgt een implantaatkaart mee als een hoog risico implantaat is toegepast bij de patiënt LET OP: Geldt zowel voor risikoklasse IIb als III implantaten (behalve hechtingen, krammen, tandheelkundige vullingen, tandheelkundige beugels, kronen, schroeven, wiggen, platen, draad, stiften, clips en connectoren)	MOR	Patiënt	Expertteam Medische hulpmiddelen	Borgen dat zorgverleners een implantaatkaart van de fabrikant mee kunnen geven waarbij de arts ook de identiteit van de patiënt op de implantaatkaart vermeldt (met sticker of pen)
V3	Er komt een nieuwe Europese databank met voor alle partijen transparante en passende informatie: EUDAMED	MOR en IVDR	Patiënt, Zorgverlener, Lab medewerker, Onderzoeker	Expertteams Medische hulpmiddelen en IVDR's en Traceerbaarheid	Hoe krijg je juiste informatie in en uit Eudamed, vanuit onze verschillende rollen (zorgaanbieder, onderzoeker, fabrikant)
V4	circa 85% van alle IVDR's komt onder toezicht van notified bodies te vallen, terwijl dat nu hooguit 20% is	IVDR	Lab medewerkers	Expertteams IVDR's en Beschikbaarheid	Borgen dat producten met de juiste certificeringen worden geleverd Borgen dat (diagnostische) testen beschikbaar blijven voor de patiëntenzorg
V5	In-huis' gemaakte testen mogen uitsluitend gemaakt en gebruikt worden binnen de eigen zorginstelling onder strikte voorwaarden (zie artikel 5.5 IVDR): - mogen niet worden overgedragen aan een andere rechtspersoon; - er moet sprake zijn van een passend kwaliteitsmanagementsysteem; - een gelijkwaardige test is niet op de markt verkrijgbaar	IVDR	Lab medewerkers Zorgverlener	Expertteam IVDR's	Voorkomen dat dit leidt tot kostenverhogingen en beschikbaarheidsrisico's. Borgen dat er geen risico's worden gelopen en organiseren van een passend kwaliteitsmanagementsysteem. Voorkomen dat referentiefunctie academische labs in gevaar komt.
V6	Zodra een test door een fabrikant op de markt is gebracht die qua prestatieniveau gelijkwaardig is aan de in huis gemaakte test, mag een ziekenhuis niet (meer) een eigen test maken en gebruiken	IVDR	Lab medewerkers	Expertteam IVDR's	Voorkomen dat dit leidt tot kostenverhogingen en beschikbaarheidsrisico's Voorkomen dat referentiefunctie academische labs in gevaar komt
V7	Door de strengere regels voor markttoelating, meer producten die in een hogere risikoklasse gaan vallen en minder notified bodies, zal markttoelating duurer worden, een langere doorlooptijd hebben en zorgen dat fabrikanten hun portfolio gaan saneren	MOR en IVDR	Zorgverlener	Expertteam Beschikbaarheid	Voorkomen dat dit leidt tot kostenverhogingen en beschikbaarheidsrisico's

Organisatie programma MDR / IVDR

Relatie met stakeholders



Stand van zaken

- / Programma is gestart en verloopt volgens planning
- / Handvatten (waar mogelijk gevalideerd) gereed begin 2020
- / Handvatten worden beschikbaar gesteld voor gebruik in alle ziekenhuizen/zorginstellingen





**END
OF
THE
CHAIN**

Fueled
by Adara●

Zorgen

- / Vertragingen en onduidelijkheden
- / Guidances komen te laat / duiding van wetgeving ontbreekt
- / Aanwijzen van notified bodies / tijd voor (her)certificering dringt
- / Uitstel van Eudamed
- / Gebrek aan informatie om continuïteit van zorg te realiseren





Nationale ontwikkelingen

1. Wet- en regelgeving
 - Maartje van der Avert
2. CCMO
 - Anneriet Heemskerk



Nationale wetgeving

- Besluit medische hulpmiddelen
- Regeling medische hulpmiddelen
- Planning: in eerste kwartaal 2020 gereed



Update CCMO

- CCMO

Verordening Medische Hulpmiddelen EU no 2017/745 hoofdstuk VI (art 61-82): klinisch onderzoek

*Centrale
Commissie
Mensgebonden
Onderzoek*

Veldbijeenkomst, 27 november 2019

Anneriet Heemskerk
Landelijk Bureau CCMO




Stappenplan

 Artikel 62 (incl. art. 74.2)


 Artikel 74.1

 Artikel 82



\$ (artikel 74, lid 2)

bijvoorbeeld in-house products



Wat zijn de praktische gevolgen? (1)

Onder andere:

- Classificatieregels aangepast: laag → hoog risico medisch hulpmiddel
- Meer klinische data nodig voor verkrijgen CE-markering
- Verplichting fabrikant: post marketing clinical follow up (PMCF) studies
- Toetsingscriteria → geen grote wijzigingen tov huidige situatie muz hoog risico medische hulpmiddelen zonder CE markering (art 62) → toets aan gemeenschappelijke specificaties (GS) of geharmoniseerde normen
- Indieningsdossier → geen grote wijzigingen tov huidige situatie muz medische hulpmiddelen zonder CE markering (art 62) → CEP
- Notificatieplicht IGJ komt te vervallen

Wat zijn de praktische gevolgen? (2)

Onder andere:

- Start gecoördineerde multinationale beoordeling ??? Voorlopig voortzetting van de nationale indieningen
- Overgangsbepaling MDR (art.120): Studies met positief besluit < 26 mei 2020 hoeven niet opnieuw beoordeeld (conform MDR) te worden door een erkende METC/CCMO
- Vigilantie bepalingen (registratie en melden SAE en SADE) – art 80: geldt voor al het lopende en nog te starten te onderzoek vanaf 26 mei 2020

Waar werken wij ondermeer aan?

- Leidraad voor METCs
- Kennisnetwerk

contact

ccmo@ccmo.nl

information

<http://www.ccmo.nl>

Centrale

Commissie

Mensgebonden

Onderzoek



Disclaimer

No rights can be derived from this presentation. Although the CCMO takes every effort to ensure the information contained in this presentation is accurate, it cannot accept any liability for the accuracy of this information.

The views and opinions expressed in this presentation are those of the author and do not necessarily represent those of the CCMO.

The CCMO accepts no responsibility for the consequences of using this presentation contained therein.

The information in this presentation may be used if the source is quoted.

Centrale

Commissie


Mensgebonden

Onderzoek





Pauze





Ministerie van Volksgezondheid,
Welzijn en Sport



Europese ontwikkelingen

1. MDCG/CAMD
2. Eudamed
3. Notified bodies
4. Brexit

*Maartje van der Avert (VWS) en
Laura de Vries (IGJ)*



CAMD – 16/17 oktober

- Overgang van DG Grow naar DG Santé
- Nationale wetgeving
- [JAMS – Joint Action on Market Surveillance](#)
 - WP4: joint manufacturer inspections
 - WP5: Clinical Process and Resource development
- Opvolger ITF en TSG: CAMD Operational
- Working Group
- Brexit





MDCG – in vogelvlucht...

- CTS under IVDD
- Common specifications under IVDR
- Transparantie
- [Communication campaign](#)
- Volgende MDCG: 13 december 2019
- Verslagen in het [Commissieregister](#)



MDCG – Guidances onlangs gepubliceerd of in aantocht

- Implant card – juni 2019
- Person responsible for regulatory compliance – juni 2019
- SCHEER guidelines phthalates – juni 2019
- Summary of Safety and Clinical Performance – augustus 2019
- Diverse guidances over notified bodies – oktober 2019
- Guidance Qualification and classification of software – oktober 2019
- Guidance Klasse I (eind 2019 in final draft)
- Guidance Clinical / Performance Evaluation of Software (eind 2019 in final draft)
- Guidance on Cybersecurity (eind 2019 in final draft)

https://ec.europa.eu/growth/sectors/medical-devices/new-regulations/guidance_en



MDCG – Stand van zaken notified bodies

- Inmiddels 7 notified bodies aangewezen voor MDR en 2 notified bodies aangewezen voor IVDR
 - Begin november 2 Nederlandse notified bodies aangewezen
- Link: <https://ec.europa.eu/docsroom/documents/35043>



EUDAMED



MDCG – Gemeenschappelijke specificaties Annex XVI

- Verdiepende sessie georganiseerd op 22 maart 2019 (verslag beschikbaar)
- Informele consultatie afgerond, opmerkingen worden nu verwerkt
- Volgende stap: start formele adoptieprocedure
- Onderdeel adoptieprocedure: formele consultatie, hopelijk voor kerst
- Na publicatie: vervolg op eerdere verdiepende sessie
 - 28 januari 2020: groepen 1-3
 - 4 februari 2020 : groepen 4-6



MDCG – Expert Panels

- Formele deadline aanmelden expert panels is gesloten op 24 november
- Aanmelding staat nog wel open voor specifieke deelterreinen waar onvoldoende experts aangemeld zijn.
- Commissie (JRC) start komende periode met de selectie.
- [Aanmelding expert panels](#)



Brexit

- Uitstel van no-deal Brexit tot uiterlijk 31 januari 2020.
- 12 december Britse Parlementsverkiezing.
- Risico op no-deal scenario Brexit blijft onverminderd hoog.
- Overstap van SGS naar België loopt nog
- Advies: leun niet achterover maar benut de extra tijd die u nu gegund wordt.
- Etikettering: géén overgangperiode van 6 maanden na no-deal in januari



Vragen?



BORREL
&
BABBEL